

**Our records show that it is time to
Recalibrate the following Gauge(s):**

Date: _____
S/N: _____ / _____

Gauges used in hydrostatic testing of TC/D.O.T. specifications cylinders should be recalibrated every 6 months to .25% at test pressures. The below table indicates the most common calibration points for gauges used in cylinder testing. If the calibration point you require is not listed, please indicate the range in the additional row provided and they will be calibrated within the TC/D.O.T. usable test range.

Full Range (psi)	Increments (psi)	TC Usable Test Range (psi)	Calibration Points (psi)
0-800	2	111-727	200, 350, 400, 500, 600, 700
0-1100	2	111-1000	200, 350, 500, 600, 700, 800, 900, 1000
0-1500	2	111-1363	200, 350, 500, 600, 700, 800, 900, 1000, 1200, 1400
0-1500	5	275-1363	200, 350, 500, 600, 700, 800, 900, 1000, 1200, 1400
0-5000	10	555-4540	1000, 2000, 3000, 4000, 5000
0-5000	20	1120-4540	1000, 2000, 3000, 4000, 5000
0-10,000	25	1388-9088	1000, 2000, 3000, 4000, 5000, 6000, 7000, 8000, 9000
0-11,000	20	1120-10,000	1000, 2000, 3000, 4000, 5000, 6000, 7000, 8000, 9000, 10,000

Recalibration service is performed on a 3-working day turnaround after receiving the gauge.

Please provide the shipment return information on the purchase accompanying purchase order.

Pack gauge carefully for shipment with form